



Certificate of Compliance

Applicant: Metrix Instrument
 Applicant's Address: 1711 Townhurst Drive
 Houston, TX 77043

Model: Models TXR and TXA
 Project Number: 06131-10

The **Metrix Instrument Models TXR and TXA** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1			
	Radiated Emissions	Class A or B	November 11, 2005
	ANSI C63.4	Class A or B	November 11, 2005
	FCC 47 CFR Part 15	Class A or B	November 11, 2005
EN 61326-1			
	EN 61000-4-2	4kV/8kV	December 2, 2005
	EN 61000-4-3	10 V/m	February 9, 2006
	EN 61000-4-4	.5kV	February 9, 2006
	EN 61000-4-6	10 Vrms	February 13, 2006

I, Michael A. Royer, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.

Michael A. Royer

Michael A. Royer
 EMC Department Manager



Lab Code 200062-0

This report has been reviewed and accepted by Metrix Instrument. The undersigned is responsible for ensuring that the Metrix Instrument Models TXR and TXA will continue to comply with the applicable rules.

Paul Dobson

METRIX DOC NO: 1184878 REV: A
