



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1090276 **Master Contract:** 168872 (LR 49334)
Project : 1090276 (Edition 1) **Date Issued:** July 13, 2000
Issued to: Metrix Instruments Co.
1711 Townhurst Drive
Houston, TX 77043
U.S.A.

The products listed below are eligible to bear the CSA Mark shown



Issued by: Dan Livermore, C.E.T.

Signature: 

CLASS

2258 02 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations

PRODUCTS

Class I, Groups A, B, C & D; Class II, Groups E, F & G

Model 7295 Explosion Proof Accelerometer Housing

APPLICABLE REQUIREMENTS

- | | | | |
|-----------------------|-----------|---|--|
| CSA Standard C22.2 No | 0-91 | - | General Requirements - Canadian Electrical Code Part II. |
| | 30-M1986 | - | Explosion-proof Enclosures for Use in Class I Hazardous Locations. |
| | 142-M1987 | - | Process Control Equipment. |

METRIX DOC NO: 1186628 REV: A



CSA INTERNATIONAL

Supplement to Certificate of Compliance

Certificate: 1090276

Master Contract: 168872 (LR49334)

Project: 1090276

Issued to: **Metrix Instruments Co.**
1711 Townhurst Drive
Houston, TX 77043
U.S.A.

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Issued By: Dan Livermore, C.E.T.

Signature

Product Certification History

Revision	Date	Description
1090276	July 13, 2000	Model 7295 Explosion Proof Accelerometer Housing.

METRIX DOC NO: 1186628
REV: A