


Certificate of Compliance

Applicant: **Metrix Instrument Co.**
Applicants Address: 1711 Townhurst Drive, Houston, Texas 77043
Model: 1001 Current Eddy Probe
Serial Number: 3660
Project Number: 96-417

The Metrix Instrument Co. 1001 Current Eddy Probe was tested to EN55022 and EN50082-2, and Metrix Instrument Co. Engineering Specifications, utilizing the following documents, and found to be in compliance with the required criteria on the indicated test dates.

EN55022: Emissions	Class A	December 19, 1995
EN61000-4-2: ESD	Level 2 (4/8 kV)	January 16, 1996
ENV50140: Radiated Immunity	Level 3 (10 V/m)	January 16, 1996
EN61000-4-4: EFT	Level 3 (2 kV)	January 16, 1996
ENV50141 : RF Immunity	Level 2 (10 V)	January 16, 1996

I, Jeffrey A. Lenk for Professional Testing (EMI), Inc., being familiar with the European Community Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.



Jeffrey A. Lenk
President

METRIX DOC No: 1163092
REV: A