



# Certificate of Compliance

Applicant: Metrix Instruments  
 8824 Fallbrook Drive  
 Houston, TX 77064

Model: SW6000 Vibration Switch  
 Project Number: 11649-10

The **Metrix Instruments SW6000 Vibration Switch** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11: 2003, Radiated Emissions	Group 1, Class A	November 3, 2010
IEC 61000-4-3: 2002	3 V/m (1.4 to 2.0 GHz) 1 V/m (2.0 to 2.7 GHz)	November 5, 2010

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.

Jeffrey A. Lenk  
 President



Lab Code 200062-0

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the Metrix Instruments SW6000 Vibration Switch will continue to comply with the applicable rules.

Metrix Instruments Representative