



Certificate of Compliance

Applicant: Metrix Instruments
8824 Fallbrook Drive
Houston, TX 77064

Model: Model SA6350 High Temperature Accelerometer
Project Number: 11654-10

The **Metrix Instruments Model SA6350 High Temperature Accelerometer** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11: 2003, Radiated Emissions	Group 1, Class A	December 1, 2010
IEC 61000-4-3: 2002	3 V/m (1.4 to 2.0 GHz) 1 V/m (2.0 to 2.7 GHz)	November 30, 2010
PTI EMC Compliance Report 04130-10		

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.

Jeffrey A. Lenk
President



Lab Code 200062-0

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the Metrix Instruments Model SA6350 High Temperature Accelerometer will continue to comply with the applicable rules.

Metrix Instruments Representative