



# Certificate of Compliance

Applicant: Metrix Instruments  
 8824 Fallbrook Drive  
 Houston, TX 77064

Model: SA6200A  
 Project Number: 11647-10

The **Metrix Instruments SA6200A** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11: 2003, Radiated Emissions	Group 1, Class A	November 1, 2010
IEC 61000-4-3: 2002	3 V/m (1.4 to 2.0 GHz) 1 V/m (2.0 to 2.7 GHz)	November 30, 2010

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.



Lab Code 200062-0

Jeffrey A. Lenk  
 President

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the Metrix Instruments SA6200A will continue to comply with the applicable rules.

Metrix Instruments Representative

METRIX DOC No: 1159091 REV: A
----------------------------------



# Certificate of Compliance

Applicant: Metrix Instruments  
 Applicant's Address: 1711 Townhurst Drive  
 Houston, TX 77043-2899  
 Model: Model SA6200A  
 Serial Number: NA  
 Project Number: 03283-10


The **Metrix Instruments, Model SA6200A** was tested to EN61326-1, Amendment 1: 1998, Annex A. The Model SA6200A was found to be in compliance with the required criteria on the indicated test dates.

EN 61326-1: 1998


Radiated Emissions	Class A	February 5, 2003
EN 61000-4-2: 1995	4kV/8kV	April 17, 2003
EN 61000-4-3: 1995	10 V/m	February 27, 2003
EN 61000-4-4: 1995	.5kV/1kV	February 26, 2003
EN 61000-4-6: 1996	3 Vrms	February 26, 2003
EN 61000-4-8: 1994	30 A/m	February 26, 2003

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.



  
 Jeffrey A. Lenk  
 President

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the **Metrix Instruments, Model SA6200A** will continue to comply with the applicable rules.



METRIX DOC No: 1159091  
 REV: A