



Certificate of Compliance

Applicant: Metrix Instruments
 8824 Fallbrook Drive
 Houston, TX 77064

Model: Model ST5484E Loop-powered Vibration Transmitter
 Project Number: 11653-10

The **Metrix Instruments Model ST5484E Loop-powered Vibration Transmitter** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11: 2003, Radiated Emissions	Group 1, Class A	November 1, 2010
IEC 61000-4-3: 2002	3 V/m (1.4 to 2.0 GHz) 1 V/m (2.0 to 2.7 GHz)	November 5, 2010
PTI EMC Compliance Report 03120-10		

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.



Lab Code 200062-0

Jeffrey A. Lenk
 President

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the Metrix Instruments Model ST5484E Loop-powered Vibration Transmitter will continue to comply with the applicable rules.

Metrix Instruments Representative

METRIX DOC NO: 1185337 REV: A
