



Certificate of Compliance

Applicant: Metrix Instruments
 8824 Fallbrook Drive
 Houston, TX 77064

Model: Model 5500 Handheld Vibration Meter

Project Number: 11651-10

The **Metrix Instruments Model 5500 Handheld Vibration Meter** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11: 2003, Radiated Emissions	Group 1, Class A	November 3, 2010
IEC 61000-4-3: 2002	3 V/m (1.4 to 2.0 GHz) 1 V/m (2.0 to 2.7 GHz)	November 5, 2010

I, Jeffrey A. Lenk, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.



Lab Code 200062-0

Jeffrey A. Lenk
 President

This report has been reviewed and accepted by Metrix Instruments. The undersigned is responsible for ensuring that the Metrix Instruments Model 5500 Handheld Vibration Meter will continue to comply with the applicable rules.

Metrix Instruments Representative