

## **EU Declaration of Conformity MX2033 and MX2034 Digital Proximity Systems**

Manufacturer's Name	Metrix Instrument Co., L.P.
Manufacturer's Address	8824 Fallbrook Drive, Houston, TX 77064 USA 281.940.1802 www.metrixvibration.com info@metrixvibration.com
Declares that the product	Model MX2033 and MX2034 Digital Proximity Systems (This declaration of conformity is issued under the sole responsibility of Metrix Instrument)
Directive 2014/34/EU ATEX	Provision of the Directive fulfilled by the Equipment  • $(x)$     1 G, Ex ia     C T4 Ga (Ta -40°C to +85°C)  • $(x)$     3 G, Ex ec     C T4 Gc (Ta -40°C to +85°C)
Conforms to the following standards	ATEX  • EN 60079-0: 2018  • EN 60079-7: 2015+A1:2018  • EN 60079-11: 2012
EU-Type Examination	<ul> <li>Certificate</li> <li>ETL22ATEX0138X (Ex ia)</li> <li>ETL22ATEX0157X (Ex ec)</li> <li>Intertek Testing &amp; Certification Limited Cleeve Road Leatherhead Surrey, KT22 7SA, United Kingdom</li> <li>Notified Body Number: 0359</li> </ul>
Notified Body for production	SGS Fimko Oy (Notified Body Number: 0598) Helsinki, Finland
Directive 2014/30/EU	<ul> <li>EMC</li> <li>EN 61326-1:2013 (self-certified)</li> <li>EN 61326-1:2006</li> <li>Professional Testing (EMI), Inc. 1601 North A.W. Grimes Blvd., Suite B Round Rock, Texas 78665 NVLAP: Lab Code 200062-0</li> </ul>
Technical File Location	The Technical File required by this Directive is maintained by the manufacturer listed above and the address listed above.
Approved By	Lance Truong Quality Manager, Metrix Instrument Co. Houston, Texas 77064 October 12, 2022