





EU Declaration of Conformity

SM6100 Electronic Switch

Manufacturer's Name	Metrix Instrument Co., L.P.
Manufacturer's Address	8824 Fallbrook Drive, Houston, TX 77064 USA 281.940.1802 www.metrixvibration.com info@metrixvibration.com
Declares that the product	Model SM6100 Electronic Switch (This declaration of conformity is issued under the sole responsibility of Metrix Instrument)
Directive 2014/34/EU ATEX	Provision of the Directive fulfilled by the Equipment <ul style="list-style-type: none">•  II 2 G, Ex db IIB +H₂ T4 Gb
Conforms to the following standards	ATEX <ul style="list-style-type: none">• EN 60079-0:2012+A11:2013• EN 60079-1:2007- This standard has been compared with EN 60079-1:2014 (currently harmonized) and no significant changes have occurred which are applicable to this equipment
EU-Type Examination	Certificate <ul style="list-style-type: none">• LCIE02ATEX6157X• Laboratoire Central Des Industries Electriques 33 avenue du General Leclerc BP 8 F92266 Fontenay-aux-Roses cedex France Notified Body Number: 0081
Notified Body for production	SGS Fimko Oy (Notified Body Number: 0598) Helsinki, Finland
Directive 2014/30/EU	EMC <ul style="list-style-type: none">• EN 61326-1: 2013 (self-certified)• EN 61326-1: 2006• Professional Testing (EMI), Inc. 1601 North A.W. Grimes Blvd., Suite B Round Rock, Texas 78665 NVLAP: Lab Code 200062-0
Technical File Location	The Technical File required by this Directive is maintained by the manufacturer listed above and the address listed above.
Approved By	 Lance Truong Quality Manager, Metrix Instrument Co. Houston, Texas 77064 October 23, 2020