



中国国家强制性产品认证证书



证书编号：2020322315003136

认证委托人名称： 儒博控制系统制造（上海）有限公司
 认证委托人地址： 上海市嘉定区马陆镇龙盘路 568 号 7 幢
 生产者名称： 迈确仪表公司
 生产者地址： 8824 Fallbrook Dr. Houston, TX 77064 United States
 生产企业名称： 迈确仪表公司
 生产企业地址： 8824 Fallbrook Dr. Houston, TX 77064 United States

产品名称： 趋近式传感器
 系列、规格、型号： 10000 系列
 标准： GB 3836.1-2010、GB 3836.4-2010

上述产品符合强制性产品认证实施规则 CNCA-C23-01:2019 的要求，特发此证。

发证日期：2020 年 11 月 4 日 有效期至：2025 年 11 月 3 日

首次发证日期：2020 年 11 月 4 日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书的相关信息可通过国家认监委网站 www.cnca.gov.cn 查询



批 准：



上海仪器仪表自控系统检验测试所有限公司

<http://www.sitiias.com.cn>

中国·上海·漕宝路103号200233

电话：+86 21 64510844

S 0003248

METRIX DOC NO: 1904419
REV: A



CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



CERTIFICATE NO: 2020322315003136

APPLICANT: Roper Industries Manufacturing (Shanghai) Co.,Ltd.
ADDRESS: Bd.7A,No.568,Longpan Rd.Malu Jiading Shanghai
MANUFACTURER: Metrix Instrument Co., LP
ADDRESS: 8824 Fallbrook Dr. Houston, TX 77064 United States
FACTORY: Metrix Instrument Co., LP
ADDRESS: 8824 Fallbrook Dr. Houston, TX 77064 United States

PRODUCTNAME: proximately sensor
SERIES,SPECIFICATION,MODEL: 10000 series
STANDARDS: GB 3836.1-2010、GB 3836.4-2010

This is to certify that the above mentioned product(s)complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

Valid from:November 4, 2020 **Valid until:** November 3, 2025

Date of original certification:November 4, 2020

The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

This certificate is available through CNCA’s website: www.cnca.gov.cn



APPROVAL:

Xu JianPing



Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.

<http://www.sitilas.com.cn>

Building 9,103 Cao Bao Road, Shanghai 200233,China

Tel: +86 21 64510844

METRIX DOC NO: 1904419
REV: A

S 0002107



中国国家强制性产品认证证书



证书编号: 2020322315003136

附件

产品名称: 趋近式传感器

型号规格: 10000 系列

Probe Series	Probe Catalog Part Number	Extension Cable Catalog Part Number	Description
10001 - 10008; 10026; 10030; 10037 - 10042	As listed		7200-compatible proximity probes, 5mm and 8mm
10000		7402	7200-compatible extension cables for 5mm and 8mm probes
10037 - 10042	As listed		7200-compatible proximity probes, 11mm
10050		9282	7200-compatible extension cables for 11mm probes
10009; 10010; 10012 - 10015; 10017 - 10019; 10025; 10027 - 10029	As listed		3000-compatible proximity probes, .190" and .300"
10075		7386	3000-compatible extension cables for .190 and .300 probes
10101 - 10106; 10171 - 10174	MX3301XX		3300-compatible proximity probes, 5mm and 8mm
10100		MX330130	3300-compatible extension cables for 5mm and 8mm probes
10901 - 10910	MX3309XX		RAM- and NSV-compatible probes, 5.26mm
10900		MX330930	RAM- and NSV-compatible extension cables for 5.26mm probes
10201 - 10206; 10271 - 10274	MX2030		3300-series compatible probes, 5mm and 8mm
10200		MX2031	3300-compatible extension cables for 5mm and 8mm probes
10301 - 10306; 10371 - 10374	MX8030		3300-series compatible probes, 5mm and 8mm
10300		MX8031	3300-compatible extension cables for 5mm and 8mm probes
ALL		5494	High pressure feed through extension cables, All probe families

防爆标志: Ex ia IIC T3/T4 Ga

电气参数: $U_i=28V$ $I_i=138mA$ $L_i=200\mu H$ $C_i=2nF$

批准:



上海仪器仪表自控系统检验测试所有限公司