




EU/UK Declaration of Conformity

SW6000 Electronic Switch

Manufacturer's Name	Metrix Instrument Co., L.P.
Manufacturer's Address	8824 Fallbrook Drive, Houston, TX 77064 USA 281.940.1802 www.metrixvibration.com info@metrixvibration.com
Declares that the product	Model SW6000 Electronic Switch (This declaration of conformity is issued under the sole responsibility of Metrix Instrument)
Directive 2014/34/EU ATEX	Provision of the Directive fulfilled by the Equipment •  II 2 G, Ex db IIB +H₂ T5 Gb • Ta: -40°C to +85°C (without display) • Ta: -10°C to +70°C (with display)
Conforms to the following harmonized/designated standards	ATEX • EN IEC 60079-0:2018 • EN 60079-1:2014 UKEX • Approval is based on existing ATEX approval
EU-Type Examination	Certificate • CML 24ATEX1294X • CML B.V. Koopvaardijweg 32, 4906CV Oosterhout, The Netherlands Notified Body Number: 2776
Notified Body for production	ATEX - SGS Fimko Oy (Notified Body Number 0598) Helsinki, Finland
Directive 2014/30/EU	EMC • EN 61326-1: 2013 (self-certified) • EN 61326-1: 2006 • Professional Testing (EMI), Inc. 1601 North A.W. Grimes Blvd., Suite B Round Rock, Texas 78665 NVLAP: Lab Code 200062-0
Technical File Location	The Technical File required by this Directive is maintained by the manufacturer listed above and the address listed above.
Approved By	 Lance Truong Quality Manager, Metrix Instrument Co. Houston, Texas 77064 July 29, 2025