



Certificate of Compliance

Applicant: Metrix
 Applicant's Address: 8824 Fallbrook
 Houston, TX 77064

Model: PRO 440
 Serial Number: Proto
 Project Number: 08249-10

The **Metrix, PRO 440** was tested utilizing the following documents and found to be in compliance with the required criteria on the indicated test date.

EN 61326-1:2006		
CISPR 11		
Conducted Emissions	Group 1, Class A	April 25, 2008
Radiated Emissions	Group 1, Class A	January 21, 2008
EN 61000-3-2	Class A	May 1, 2008
EN 61000-3-3		May 1, 2008
FCC 47 CFR Part 15	Class A	January 21, 2008
EN 61326-1:2006		
EN 61000-4-2	4 kV/8 kV	April 28, 2008
EN 61000-4-3	1 V/m, 3 V/m, 10 V/m	June 2, 2008
EN 61000-4-4	1 kV/2 kV	May 2, 2008
EN 61000-4-5	1 kV/2 kV	October 23, 2008
EN 61000-4-6	3 Vrms	May 2, 2008
EN 61000-4-8	30 A/m	April 2, 2008
EN 61000-4-11	100% - 1p, 100% 0.5 each polarity	January 26, 2009

I, Jason Anderson, for Professional Testing (EMI), Inc., being familiar with the Electromagnetic Compatibility rules and test procedures, have reviewed the test setup, measured data and this report. I believe them to be true and accurate.

Jason Anderson
 Director of Testing Services

Lab Code 200062-0

This report has been reviewed and accepted by Metrix. The undersigned is responsible for ensuring that the Metrix PRO 440 will continue to comply with the applicable rules.